

Analysis of Anisole (Methoxybenzene) by HPLC-UV

Column: [Waters XBridge C18 2.5µm 2.1 X 50mm Column \(Ctrl + Click to follow link\)](#)

Elution Type: Isocratic

Elution A: Water

Elution B: MeOH


Gradient Profile:

Step No.	Time (min)	Pct A	Pct B
1	0	60	40

Flow Rate: 1.85 mL/min

Col. Temp: Ambient

Detection: [UV-Vis Abs.-Diode Array \(PDA\) @ 254 nm \(22°C\) \(Ctrl + Click to follow link\)](#)



1: 1,2-Dimethoxybenzene

2: 1,2,4-Trimethoxybenzene

3: 1,4-Dimethoxybenzene

4: 1,3-Dimethoxybenzene

5: 1,3,5-Trimethoxybenzene